



## cellSens Dimension

For Compatibility: See Software Provider

Illumination System/Accessory	Recognised in cellSens as:	Required cellSens version	USB Control	TTL Control (via external hardware)	Analog Control (via external hardware)	Further software information:	CoolLED Driver
<b>pE-2</b>	precisExcite	1.14 or later	Individual Wavelength control	Individual Wavelength control	N/A	<a href="#">Click for link Olympus cellSens website</a>  <a href="#">Click for link to CoolLED Easy Integration Guides</a>	Windows XP onwards. 32-bit & 64-bit <a href="#">Click for link to CoolLED driver</a>
<b>pE-300<sup>white</sup></b>	pE-300 white	2.3 or later	Individual Wavelength control	Global control	N/A		
<b>pE-300<sup>ultra</sup></b>	pE-300 ultra	2.3 or later	Individual Wavelength control	Global, Individual and sequential Wavelength control	N/A		
<b>pE-340<sup>fura</sup></b>	pE-340 fura	2.3 or later	Individual Wavelength control	Global, Individual and sequential Wavelength control	N/A		
<b>pE-4000</b>	pE-4000	1.14 or later	Individual Wavelength control	Individual Wavelength control	Individual Wavelength control		
<b>pE-6501 pE-6501-8 pE-6501-8-FR</b>	NI6501	1.17 or later	N/A	N/A	N/A	<a href="#">pE-6501 Guide</a>  <a href="#">pE-6501-8 Guide</a>	