



America

CERTIFICATE

No. U8 123636 0003 Rev. 01

Holder of Certificate: CoolLED Limited
26 Focus Way
Andover, Hampshire SP10 5NY
UNITED KINGDOM

Certification Mark:



Product: Laboratory Equipment
LED Light Source
Brand Name: CoolLED

Model(s): pE-400 SB, pE-400 MB, pE-400^{max} SB, pE-400^{max} MB
Parameters: Rated Voltage: 12VDC
Rated current: 8.5A
Protection Class: Class III

Tested according to: UL 61010-1:2012/R:2023-06
CSA C22.2 No. 61010-1:2012/U3:2023-06

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing, Certification, Validation and Verification Regulations (TCVVR)". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: 071-75960341T-001

Date, 2024-07-31

(Matthew Emery)